

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

FEB 08 2010

2/8/10

130288

DataMaster cdm S/N

13028

2/15/2010

Supervisor/Operator Performing the Verification Procedure:

Name Chris Nelson

ID# 6472

Date 2/4/10

2/8/10
Camp

A

Agency AST

Phone # 675-4398

Instrument Location Aniak

B

Alco S/N 56799

Target Value .084

High Pressure 12

Alco Test Values

.084
.083
1st Alco

.085
.086
2nd Alco

2/8/10
Camp

Signature

Chris Nelson

(OVER)

BMB
2/10/2010

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.

(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.

(3) I am the Scientific Director of the State Breath Alcohol Program.

(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.

(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 10th day of Feb, 2010.

Carolyn M. Noland
(Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130288 *bmb*
2/9/2010
130288

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130288

FEBRUARY 05, 2010

OPERATOR'S NAME:

NELSON/CHRIS/J

OPERATOR'S NUMBER: 6472

SUBJECT'S LAST NAME:

TEST

SUBJECT'S FIRST NAME/MI :

TEST

O.L. #: 0000

DEPT/AGENCY: AN10

CASE/REPORT: 10-000000

TEST TYPE: T

ALCO TARGET VALUE: .084

ALCO S/N: 56799

--- BREATH ANALYSIS ---

.084 ADJUSTED FOR 29.72 in

ALCO TARGET .083 08:29

BLANK TEST .000 08:30

INTERNAL STANDARD VERIFIED 08:30

ALCO TV 29.72 in .086 08:31

BLANK TEST .000 08:31

SUBJECT SAMPLE .000 08:32

BLANK TEST .000 08:33

ALCO TV 29.73 in .085 08:33

BLANK TEST .000 08:34

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130288

FEBRUARY 05, 2010

TIME 08:25

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 29.72 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMNOPQRSTUVWXYZ\]^_`abcde*ghijklmno
pqrstuvwxyz{|}~#